



Adesh Institute of Medical Sciences & Research

Barnala Road, Bathinda (Pb.)

PH: Reception 0164-2860000, Emergency: 2860100

e-Mail : adeshfoundation@hotmail.com

Department Microbiology

DATE:-20/05/06

Swabs Taken From O.Ts

1. O.Ts TABLE :- NO GROWTH
2. O.Ts FLOOR:- NO GROWTH
3. O.Ts WALL:- NO GROWTH
4. O.Ts SHELF:- NO GROWTH
5. FORMALINE CHAMBER:- NO GROWTH
6. SUCTION APP.:- NO GROWTH
7. INSTRUMENT TROLLEY:- NO GROWTH
8. LIGHT :- NO GROWTH


LAB Tech.


DR. Deepak Arora

✓ Note : This Report not meant for Medico-legal purpose.



Adesh Institute of Medical Sciences & Research

Barnala Road, Bathinda (Pb.)

PH: Office 0164-2742900, FAX: 2742901

e-Mail : adeshfoundation@hotmail.com

REPORT

POST FUMIGATION STERILIZATION EVALUATION OF O.T

FOGSTAR ULV (Fumigation Machine) of make Radiant Enterprise, Ahmedabad was demonstrated in Emergency O.T of ADESH INSTITUTE OF MEDICAL SCIENCE & RESEARCH, BATHINDA on dated 18th May 2006. After fumigation, Eight (8) Swabs were taken for culture. The culture reports enclosed herewith. Sterilization of O.T is satisfactory.

Chemical used for sterilization as per company name is INCIDUR Make HENKEL ECOLAB, GERMANY (GLYOXAL + GLUTARAL)

Referred By - **Dr. Jasbeer Kaur** (Prof. & HOD Anesthesia & All O.T)

Method of Evaluation – SWABS

Total No. Of Swabs Taken- Eight (8)

Comment

The sterilization done by Fogstar ULV with INCIDUR Disinfectant is Satisfactory & the culture report of Microbiology department enclosed herewith.

Dr. Jasbeer Kaur
Prof. & HOD Anesthesia

Dated- 20th May 2006.