

ORIGINAL

Dtd.20/11/03.

To,
Dr.Pankajbhai Shah,
Director,
Gujrat Cancer Research Institute,
Civil Hospital campus,
Ahmedabad.

Ref: Your po no-GCS/NO/327 dtd.8/10/03 & our supply of Fogstar-ULV machines & disinfectant ON 15/10/03.

Subject :-Commissioning of 2nos FOGSTAR-ULV foggers & INCIDUR for fumigation in your Operation theatre complex.

Resp Sir,


Referring to above subject we are pleased to inform you that we have successfully commissioned the two foggers & demonstrated once again Fumigation protocol with FOGSTAR-ULV fogger & incidur disinfectant by fumigating your Operation theatre complex(6nos OTs) on 8/11/03 Saturday. We have also demonstrated & shown how to use machines as well as submitted fumigation protocol in written to Mr.Parbatbhai & Ushaben (OT incharge) & their staff.

Kindly arrange to release the payment at the earliest.


Thanking you

Yours sincerely,

For Radiant Enterprise,


Parth Patel.
Prop.

2 units of Fogstar-ULV installed
Demonstrated & fumigation procedure
Completed in all 6 OTs & stores.
Satisht factory. (on: 8/11/03 Saturda.
N.B. Post. fumigation Culture
Reprot is attached herewith


(P.N.Patel)
25-11-03

HIN JPS
25/11/03

Report

Date:- 17/11/03

Post Sterilization Evaluation of 6 Operation Theater's

(1) Chemical used for Sterilization as per company name:- "INCIDUR".
"The chemical nature of which was not informed to us"

(2) Requested by : Dr. S.V.Shah, Chief of Surgical oncology.

(3) Method of evaluation : Settle Plate Method


(4) Total Number of Plates evaluated / Exposed - 150

(5) All the OTs showed -
6-7 cfu's of Staph.aureus (Multi drug resistant)
1-2 cfu's of Pseudomonas
Fungus : 2-3 cfu's-Aspergillus

Comment

(1) Sterilization of the OTs was adequately performed

(2) The presence of the above said organisms (Insignificantly presence) will not lead to post operative Infection.


Dr. Parjath N. Goswami
Professor & Head
Department of Microbiology